

NAD-Malate Dehydrogenase Microplate Assay Kit

Catalog Number: AS0199



Detection and Quantification of NAD-Malate Dehydrogenase Activity in Serum, Plasma, Tissue extracts, Cell lysate, Cell culture media and Other biological fluids Samples.

This instruction must be read in its entirety before using this product.

For research use only. Not for use in diagnostic procedures.

Contact information

Tel: 1-301-446-2499

Fax: 1-301-446-2413

Email: tech@signalwayantibody.com

Web: www.sabbiotech.com

I. INTRODUCTION

Malate dehydrogenase (EC 1.1.1.37) (MDH) is an enzyme that reversibly catalyzes the oxidation of malate to oxaloacetate using the reduction of NAD⁺ to NADH. This reaction is part of many metabolic pathways, including the citric acid cycle.

NAD-Malate Dehydrogenase Assay Kit is a sensitive assay for determining

NAD-Malate Dehydrogenase activity in various samples. NAD-Malate Dehydrogenase activity is determined by NADH decomposition rate. The reaction products can be measured at a colorimetric readout at 340 nm.

II、KIT COMPONENTS

Component	Volume	Storage
96-Well Microplate	1 plate	
Assay Buffer	30 ml x 4	4 °C
Reaction Buffer	20 ml x 1	4 °C
Substrate	Powder x 1	-20 °C
Standard	Powder x 1	-20 °C
Technical Manual	1 Manual	

Note:

Substrate: add 19 ml Reaction Buffer to dissolve before use.

Standard: add 1 ml distilled water to dissolve before use; then add 0.2 ml into 0.8 ml distilled water, the concentration will be 400 $\mu\text{mol/L}$.

III.MATERIALS REQUIRED BUT NOT PROVIDED

1. Microplate reader to read absorbance at 340 nm
2. Distilled water
3. Pipettor, multi-channel pipettor
4. Pipette tips
5. Mortar
6. Centrifuge
7. Timer
8. Ice

IV.SAMPLE PREPARATION

1.For cell and bacteria samples

Collect cell or bacteria into centrifuge tube, discard the supernatant after centrifugation, add 1 ml Assay buffer for 5×10^6 cell or bacteria, sonicate (with power 20%, sonicate 3s, interval 10s, repeat 30 times); centrifuged at 10,000g 4 °C for 15 minutes, take the supernatant into a new centrifuge tube and keep it on ice for detection.

2.For tissue samples

Weigh out 0.1 g tissue, homogenize with 1 ml Assay buffer on ice, centrifuged at 10,000g 4 °C for 10 minutes, take the supernatant into a new centrifuge tube and keep it on ice for detection.

3. For liquid samples

Detect directly.

V. ASSAY PROCEDURE

Warm all reagents to room temperature before use.

Add following reagents into the microcentrifuge

Reagent	Sample	Standard	Blank
Standard	--	200 μ l	--
Distilled water	--	--	200 μ l
Substrate	190 μ l	--	--
Sample	10 μ l	--	--

Mix, measured at 340 nm and record the absorbance of 10th second and 130th second.

Note:

- 1) Perform 2-fold serial dilutions of the top standards to make the standard curve.
- 2) For unknown samples, we recommend doing a pilot experiment & testing several doses to ensure the readings are within the standard curve range. If the enzyme activity is lower, please add more sample into the reaction system; or increase the reaction time; if the enzyme activity is higher, please dilute the sample, or decrease the reaction time.

VI.CALCULATION

Unit Definition: One unit of NAD-Malate Dehydrogenase activity is defined as the enzyme reduces 1 nmol of NADH per minute.

1.According to the protein concentration of sample

$$\begin{aligned}\text{NAD-MDH (U/mg)} &= (\text{CStandard} \times \text{VStandard}) \times (\text{ODSample}(10\text{S}) - \\ &\quad \text{ODSample}(130\text{S})) / (\text{ODStandard} - \text{ODBlank}) / (\text{VSample} \times \\ &\quad \text{CProtein}) / \text{T} \\ &= 4000 \times (\text{ODSample}(10\text{S}) - \text{ODSample}(130\text{S})) / \\ &\quad (\text{ODStandard} - \text{ODBlank}) / \text{CProtein}\end{aligned}$$

2.According to the weight of sample

$$\begin{aligned}\text{NAD-MDH (U/g)} &= (\text{CStandard} \times \text{VStandard}) \times (\text{ODSample}(10\text{S}) - \\ &\quad \text{ODSample}(130\text{S})) / (\text{ODStandard} - \text{ODBlank}) / (\text{VSample} \times \text{W} / \\ &\quad \text{VAssay}) / \text{T} \\ &= 4000 \times (\text{ODSample}(10\text{S}) - \text{ODSample}(130\text{S})) / (\text{ODStandard} - \\ &\quad \text{ODBlank}) / \text{W}\end{aligned}$$

3.According to the quantity of cells or bacteria

$$\begin{aligned}\text{NAD-MDH (U}/10^4) &= (\text{CStandard} \times \text{VStandard}) \times (\text{ODSample}(10\text{S}) - \\ &\quad \text{ODSample}(130\text{S})) / (\text{ODStandard} - \text{ODBlank}) / (\text{VSample} \times \\ &\quad \text{N} / \text{VAssay}) / \text{T} \\ &= 4000 \times (\text{ODSample}(10\text{S}) - \text{ODSample}(130\text{S})) / \\ &\quad (\text{ODStandard} - \text{ODBlank}) / \text{N}\end{aligned}$$

4. According to the volume of sample

$$\begin{aligned} \text{NAD-MDH (U/ml)} &= (\text{C}_{\text{Standard}} \times \text{V}_{\text{Standard}}) \times (\text{OD}_{\text{Sample}(10\text{S})} - \\ &\quad \text{OD}_{\text{Sample}(130\text{S})}) / (\text{OD}_{\text{Standard}} - \text{OD}_{\text{Blank}}) / \text{V}_{\text{Sample}} / \text{T} \\ &= 4000 \times (\text{OD}_{\text{Sample}(10\text{S})} - \text{OD}_{\text{Sample}(130\text{S})}) / \\ &\quad (\text{OD}_{\text{Standard}} - \text{OD}_{\text{Blank}}) \end{aligned}$$

$\text{C}_{\text{Standard}}$: the standard concentration, $400 \mu\text{mol/L} = 400 \text{ nmol/ml}$;

$\text{V}_{\text{Standard}}$: the volume of standard, $200 \mu\text{l} = 0.2 \text{ ml}$;

$\text{C}_{\text{Protein}}$: the protein concentration, mg/ml ;

W : the weight of sample, g ;

N : the quantity of cell or bacteria, $\text{N} \times 10^4$;


V_{Sample} : the volume of sample, 0.01 ml ;

V_{Assay} : the volume of Assay buffer, 1 ml ;

T : the reaction time, 2 minutes .

VII.TYPICAL DATA

The standard curve is for demonstration only. A standard curve must be run with each assay.



Detection Range: 4 µmol/L - 400 µmol/L

VIII. TECHNICAL SUPPORT

For troubleshooting, information or assistance, please go online to www.sabbiotech.cn or contact us at techcn@signalwayantibody.com