

## Human Transgelin (TAGLN) ELISA Kit

Catalog No: #EK6896



Package Size: #EK6896-1 48T #EK6896-2 96T

Orders: [order@signalwayantibody.com](mailto:order@signalwayantibody.com)Support: [tech@signalwayantibody.com](mailto:tech@signalwayantibody.com)

## Description

Product Name	Human Transgelin (TAGLN) ELISA Kit
Applications	ELISA
Species Reactivity	Human (Homo sapiens)
Other Names	DKFZp686B01212; DKFZp686P11128; SM22; SMCC; TAGLN1; WS3-10; SM22-alpha[transgelin variant 2
Accession No.	Q01995
Storage	<p>The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition.</p> <p>The loss rate was determined by accelerated thermal degradation test. Keep the kit at 37C for 4 and 7 days, and compare O.D.values of the kit kept at 37C with that of at recommended temperature. (referring from China Biological Products Standard, which was calculated by the Arrhenius equation. For ELISA kit, 4 days storage at 37C can be considered as 6 months at 2 - 8C, which means 7 days at 37C equaling 12 months at 2 - 8C).</p>

## Application Details

Detect Range:Request Information

Sensitivity:Request Information

Sample Type:Serum, Plasma, Other biological fluids

Sample Volume: 1-200 µL

Assay Time:1-4.5h

Detection wavelength:450 nm

## Product Description

Detection Method:SandwichTest principle:This assay employs a two-site sandwich ELISA to quantitate TAGLN in samples. An antibody specific for TAGLN has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and anyTAGLN present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for TAGLN is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of TAGLN bound in the initial step. The color development is stopped and the intensity of the color is measured.

## Published Papers

el at., Assessing serum levels of SM22α as a new biomarker for patients with aortic aneurysm/dissection. In PLoS One on 2022 Mar 31 by Ning Zhang, Ying-Ying Wang,et al..PMID:35358189, , (2022)

[PMID:35358189](https://pubmed.ncbi.nlm.nih.gov/35358189/)

---

Note: This product is for in vitro research use only and is not intended for use in humans or animals.